


Lot-to-Lot Reproducibility ZORBAX SB-C8

Application
Technical
Robert Ricker



Conditions:
ZORBAX SB-C8, Test of Packing Lots
Mobile Phase: 1:1 ACN:0.1mM Na₂HPO₄
1 mL/min, Temp. Ambient, Detect. UV (254nm)

Highlights

- Each lot of ZORBAX StableBond packing is tested against a rigid set of quality assurance specifications before column production.
- Each column must then pass a number of quality control specifications.



Agilent Technologies

*Robert Ricker is an application chemist
based at Agilent Technologies, Wilmington,
Delaware.*

For more information on our products and
services, visit our website at:
www.agilent.com/chem

Copyright© 2002 Agilent Technologies, Inc.
All Rights Reserved. Reproduction,
adaptation or translation without prior
written permission is prohibited, except as
allowed under the copyright laws.

Agilent shall not be liable for errors
contained herein or for incidental or
consequential damages in connection with
the furnishing, performance, or use of this
material.

Information, descriptions, and specifications
in this publication are subject to change
without notice.

Printed in the USA
April 25, 2002
5988-6473EN



Agilent Technologies