


Agilent ZORBAX Eclipse Plus Phenyl-Hexyl Rapid Resolution Threaded Column



Data Sheet

General Description

Eclipse Plus Phenyl-Hexyl columns are designed for superior peak shape with basic compounds, and deliver high efficiency and excellent peak shape with all sample types. Eclipse Plus Phenyl-Hexyl is especially useful for the separation of acidic, basic, and other highly polar compounds by reverse-phase liquid chromatography. Eclipse Plus Phenyl-Hexyl packing is made by first chemically bonding a dense monolayer of dimethylphenylhexylsilane stationary phase to a specially prepared, improved ultra-high purity (>99.995% SiO₂) ZORBAX Rx-SIL porous silica support. This special silica support (Type B) is designed to reduce or eliminate strong adsorption of basic and highly polar compounds. The bonded-phase packing is then doubly endcapped using proprietary reagents and procedures to obtain maximum deactivation of the silica surface. Eclipse Plus Phenyl-Hexyl columns can be used for acidic and neutral samples, but are especially suited for separating basic compounds that produce poor peak shapes on other columns. These columns can be used for a wide range of applications and over a pH range of 2 to 8, accommodating most popular mobile phases.

The uniform, spherical Eclipse Plus Phenyl-Hexyl particles are based on an improved ZORBAX Rx-SIL support that has a nominal surface area of 160 m²/g and a controlled pore size of 95 Å. Columns are loaded to a stable, uniform bed density using a proprietary high-pressure slurry-loading technique to give maximum column efficiency.

Column Characteristics

A typical quality control (QC) test chromatogram for a 4.6 mm × 30 mm column is shown in Figure 1. The actual QC test and performance of your column is described in the Column Performance Report enclosed with your column.


Safety Considerations

- All points of connection in liquid chromatographic systems are potential sources of leaks. Users of liquid chromatographic equipment should be aware of the toxicity or flammability of their mobile phases.
- These Rapid Resolution threaded columns have been tested to a very high pressure to ensure safe opera-

Figure 1
ZORBAX Eclipse Plus Rapid Resolution Phenyl-Hexyl Chromatogram

Operating Conditions

Column:	Eclipse Plus Phenyl-Hexyl, 4.6 mm × 30 mm, (3.5 μm)
Mobile phase:	60% Acetonitrile/40% water
Flow rate:	1.0 mL/min
Temperature:	Ambient
Injection:	5 μL
Detector:	UV 254 nm
Peak identity:	1. Uracil 5 μg/mL
	2. Phenol 200 μg/mL
	3. 4-Chloronitrobenzene 25 μg/mL
	4. Toluene 850 μg/mL



tion on a variety of LC instruments. The 2.1-mm and 3.0-mm id columns will support 20,000 psi (1,300 bar) operation and 4.6-mm id columns will support 16,000 psi (1,000 bar) operation. Opening columns may compromise this pressure limit. Columns have not been tested above 600 bar.

- Because of the small particle size, dry ZORBAX packings are respirable. Columns should only be opened in a well-ventilated area.

Operational Guidelines

- The direction of flow is marked on the column.
- While it is not harmful to the column, reverse flow should be avoided except to attempt removal of inlet blockage (see "Column Care").
- A new column contains a mixture of acetonitrile and water. Initially, care should be taken not to pass any mobile phase through the column that might cause a precipitate.



Agilent Technologies

- Eclipse Plus Phenyl-Hexyl is compatible with water and all common organic solvents.
- The use of a guard column is recommended to protect the Eclipse Plus Phenyl-Hexyl column and extend its useful lifetime.
- Avoid use of this column below pH 2 or above pH 8.
- Maximum operating pressure is 600 bar (9000 psi).
- Maximum operation temperature is 60 °C.

NOTE: Eclipse Plus Phenyl-Hexyl columns are designed for high stability over a wide pH range. However, all silica-based packings have some solubility in pH > 6 aqueous mobile phases. Therefore, when using silica-based columns under conditions of pH > 6, maximum column lifetime is obtained by operation at low temperatures (< 40 °C) using low buffer concentrations in the range of 10 to 20 mM. Column stability at pH > 6 is also enhanced by avoiding phosphate and carbonate buffers [ref.: H. A. Claessens, M. A. van Straten and J. J. Kirkland, *J. Chromatogr. (A)*, 728 (1996) 259].

- Columns should not be maintained at a neutral or elevated pH, or at elevated temperatures, when not in use.

Mobile Phase Selection

The phenyl-hexyl bonded phase can have unique selectivity that is strongly influenced by the choice of organic modifier in the mobile phase. The columns are best used with methanol/water or acetonitrile/water mobile phases, but methanol/water mobile phases will enhance interactions with aromatic compounds.

Due to the relatively high viscosity of recommended mobile phases, increased efficiency can be achieved with the use of column temperatures in the range of 40 to 60 °C; however, best column lifetime is achieved with operation at < 40 °C. Gradient-elution techniques for this packing often use 5% methanol or acetonitrile as the initial solvent and 100% methanol or acetonitrile as the final solvent. Additional information on solvent selection may be found in chapters 6 and 7, *Introduction to Modern Liquid Chromatography*, Second Edition, L. R. Snyder and J. J. Kirkland, (John Wiley & Sons, 1979), and chapters 6, 7, and 8, *Practical HPLC Method Development*, Second Edition, L. R. Snyder, J. L. Glajch, and J. J. Kirkland, (John Wiley & Sons, 1997).

Applications

Eclipse Plus Phenyl-Hexyl can be used with basic, neutral, or acidic compounds. This column is an excellent choice when an alternate selectivity to the Eclipse Plus C18 is needed to enhance resolution. Resolution is especially enhanced with samples containing aromatic compounds in methanol-containing mobile phases. The most dramatic changes in selectivity will be seen with polyaromatic compounds.

Column Care

The inlet frit on these columns has a nominal porosity of 0.3 µm. Samples that contain particulate matter larger than 0.3 µm will plug the column inlet frit. Eclipse Plus

guard columns and a hardware kit are recommended for use with such samples.

If solvent flow appears to be restricted (high column back-pressure), check first to see that solvent flow is unobstructed up to the column inlet. If the column has the restriction, there may be particulate matter on the inlet frit. An attempt should be made to remove any inlet debris by back-flushing 25 to 30 mL of mobile phase through the column. If this fails to return the column to near its original operating pressure, consider replacing the column.

To remove strongly retained materials from the column, flush the column with stronger (less polar) solvents. Solvents such as methanol, acetonitrile, or a 95%/5% mixture of dichloromethane and methanol should remove most highly retained compounds. In extreme cases, dimethyl sulfoxide or dimethylformamide at low flow rates may also be used for this purpose. When switching between solvents with vastly different polarities, it may be necessary to first purge the column with a mutually miscible solvent such as isopropanol.

Storage Recommendations

Long-term storage of silica-based, bonded-phase columns should be in a pure organic solvent, preferably an aprotic liquid such as 100% acetonitrile. If the column has been previously used with a buffered mobile phase, the buffer should first be removed by purging the column with 20–30 column volumes of a 50/50 mixture of methanol or acetonitrile and water, followed by 20–30 column volumes of the pure solvent. Before storing the column, the end-fittings should be tightly capped with end-plugs to prevent the packing from drying out.

Columns may be safely stored for short periods in most mobile phases. However, to protect equipment, it is desirable to remove salts from the instrument and column by purging the column with the same mobile phase without the buffer (for example, using 60/40 ACN/H₂O to remove a 60/40 ACN/0.02 M phosphate buffered mobile phase). Re-equilibration is rapid with the original mobile phase when using this approach, and any danger of corrosion from the salts is eliminated.

Agilent Ordering Information

For more information on our products, visit our Web site at www.agilent.com/chem/supplies.

For Technical Support in the U.S. and Canada, call 1-800-227-9770 or call your local Agilent sales office.

Agilent shall not be liable for errors contained herein or for incidental or consequential damages in connection with the furnishing, performance, or use of this material.

Information, descriptions, and specifications in this publication are subject to change without notice.

© Agilent Technologies, Inc. 2008

Printed in the USA
June 24, 2008
Part No. 820118-012



Agilent Technologies